

KIMBERLY-CLARK* SAFESKIN*
Purple Nitrile-Xtra* Exam Gloves



PRODUCT INFORMATION

Product Numbers	Size	Catalog Numbers
	X-Small (XS)	55090
	Small (S)	50601
	Medium (M)	50602
	Large (L)	50603
	X-Large(XL)	50604

Material: Acrylonitrile Butadiene, latex-free, powder-free
 Tested for use with ten chemotherapy drugs
Design: Ambidextrous, beaded cuff, with textured fingertips
Packaging: 50 gloves per dispenser box. 10 Dispenser boxes per case. Shipping case and inner packaging labeled with product name, size, lot number and catalog number.

This glove meets the emergency medical examination glove requirements of the NFPA 1999, *Standard On Protective Clothing for Medical Emergency Operations*, 2003 Edition.

PHYSICAL PROPERTIES

Characteristics	Value	Test Method
Freedom from holes	1.5 AQL ¹	ASTM D 5151

¹AQL as defined per ISO 2859-1 for sampling by attributes

Tensile Properties	Tensile Strength	Ultimate Elongation	
- Before Aging	21 MPa, nominal	550% nominal	ASTM D 412 and ASTM D 573
- After Accelerated Aging	21 MPa, nominal	500%, nominal	

Dimensional	Measured Point	mm		mil		
- Nominal Thickness	Middle Finger	0.15		5.9		ASTM D 3767 and D 6319
	Palm	0.12		4.7		
	Cuff	0.09		3.5		
- Nominal Length		305 mm/ 12 inches				

Palm Widths						
- Nominal Width (mm)	X-Small	Small	Medium	Large	X-Large	ASTM D 3767 and D 6319
	70	80	95	110	120	

(800) 255-6401 • www.kc-safety.com

*Registered Trademark or Trademark of Kimberly-Clark Worldwide, Inc.
 6-20-06